

# Certificate of Conformity CoC

**Electrosuisse CERTIFEL CoC**  
(Type testing certification scheme)

Certificate number

**HG-3082**

page 1 of 1

Product	Air purification unit	
Applicant	MedicCleanAir Melkmarkt 35, BE-2000 Antwerpen	
Manufacturer	MedicCleanAir Melkmarkt 35, BE-2000 Antwerpen	
Manufactured at/Factory	ID 101831	
Trade-mark	MedicCleanAir	
Type/Model	PRO 100, PRO 110 ISO 120, ISO 210, ISO 520	
Ratings, characteristics	PRO 100, PRO 110: 220-240 V, 50-60 Hz, 85 W ISO 120, ISO 210: 220 V, 60 Hz, 85 W Class I, IPX0  ISO 520: 220-240 V, 50-60 Hz, 175 W 220 V, 60 Hz, 175 W Class I, IPX0	
Standards	EN 60601-1-2:07 IEC 60601-1-2(ed.3)	
Additional information	EMC requirements of class 1 medical electrical equipment according to following regulation:	
EC-Directives Commission Regulations	93/42/EC	Medical devices
Swiss regulations	SR 812/213	Medizinprodukteverordnung (Ordinance concerning Medical Devices)

The products submitted by the applicant was type tested in accordance with the requirements of the mentioned standards. Electrosuisse certifies the conformity of the product with the standards. Presumption of conformity with the mentioned European directives and Swiss regulations can be assumed. Assurance of conformity during production will be in responsibility of the manufacturer. This certificate has been issued under the presumption and conditional on the fact that the applicant holds all necessary legal rights with regard to the product presented for testing and certification.

Results of testing are shown in report number: 12-HG-0187.10 + .11 + .12 + .13

 **Electrosuisse**  
**Swiss Certification Body**

Martin Plüss  
Product Certification




SCESp 036

Fehraltorf, 2014-10-06